

Item number 1999903451

Classic

Examination gloves, Abena Classic, M, nature, latex, powder free, single use

- ✓ Full textured
- ✓ High grip
- ✓ Elastic
- ✓ Good fit
- ✓ Good sensitivity

**Product Description**

Examination gloves with textured surface. Suitable for handling foods with no fat content and for brief cleaning tasks without strong chemicals.

**HIGH GRIP****EN EN
455 420**

EN374-5:2016

374-1:2016/
Type B**AQL
1.5****Specifications**

Base name	Examination gloves
Brand	Abena
Sub-brand	Classic
Size	M
Sub-color	Nature
Properties	Full textured, high grip, rolled cuff
Features	Powder free
Single or multiple use	Single use
Material	Latex
Technical ingredients	Latex, ZDBC, ZDEC, Zinc oxide, Sulfur, Titanium Dioxide, Calcium Nitrate, Polymer, Freesil P2, WAX
Length/depth	240 mm
Width	95 mm
Certifications	CAT III, CE, Food contact materials
CE Category (Personal Protective Equipment)	CAT III
CE Class (Medical Devices)	Class I
Product or test standards	AQL 1.5, EN 374-1:2016 Type B KST, EN 374-5:2016 VIRUS, EN 420, EN 455
Directives, regulations and acts	1935/2004/EU
Safety Instructions and Warnings	NB: If users or patients are hypersensitive to the material, wear vinyl or nitrile gloves. Contains latex proteins.
Shelf Life	3 years
Shelf Life, opened	3 years
Storage Instructions	Store dry, clean and at room temperature.
Product Disposal Instructions	To be disposed of with ordinary household waste. If contaminated, dispose of as clinical waste.
Packaging Disposal Instructions	Can be recycled or incinerated.

Instructions for use/application

Examine the gloves for flaws and defects before use.

Packaging data

Unit	Contains	Length	Width	Height	EAN
cII	10 pck	36 cm	24.5 cm	22.5 cm	5703538445316
pck	100 pcs	21.5 cm	12 cm	7 cm	5703538445309
pcs					



The glass fork symbol guarantees that products have been tested in accordance with European legislation and approved for food contact. The symbol is mandatory on products used for food contact.



The CE mark guarantees that a product is safe to use and complies with all safety precautions. CE stands for Conformité Européenne (European Conformity) and is mainly found on electronic equipment, safety equipment, construction products and medical equipment.



Third-party type approval is required for all Category III personal protective equipment (PPE) products. The showing of the CE logo for chemical protective gloves requires that tests are carried out in accordance with test standards specified in EN ISO 374-1: 2016 + A1: 2018 - such as EN 16523-1: 2015 + A1: 2018 to determine the resistance to the permeation of chemicals. The results of this test determine the relevant pictogram symbols that can be used on the packaging and labeling.



The glove is 240 mm long.



The product contains latex.



The glove has a high grip.



The glove has a fully textured surface.



EN 455 consists of four standards that gloves must be tested against in order to be a medical glove. This standard specifies the requirements for disposable medical gloves.



The standard specifies the requirements for protective gloves in relation to general requirements and test methods.

EN374-5:2016



The standard specifies the requirements for protective gloves against hazardous chemicals and microorganisms. This part of the standard describes the glove's resistance to bacteria, fungi and viruses.

374-1:2016/
Type B



The standard specifies the requirements for protective gloves against chemicals and microorganisms. Type B has been tested for a permeation time of at least 30 minutes with the chemicals sodium hydroxide 40%, hydrochloric acid 40% and formaldehyde 37%.