



Item number 290415



Examination gloves, ABENA Classic Protect, XL, white, nitrile, powder free

- ✓ Elastic
- ✓ Fits to hand
- ✓ No latex proteins



CAT
III

Product Description

Elastic examination glove that fits to the shape of the hand during use. Can be used for a variety of tasks such as cleaning, care, food handling, laboratory, medication handling, cytostatics and hormone creams. Contains no latex proteins therefore safe for use in the event of latex allergy.



MEDIUM
GRIP



EN374-5:2016



374-1:2016/
type B



EN EN AQL
420 455 1.0

EN 421:2010



Specifications

Base name	Examination gloves
Brand	ABENA
Sub-brand	Classic Protect
Size	XL
Color	White
Properties	Finger textured, latex free, medium grip, rolled cuff
Features	Powder free
Single or multiple use	Single use
Material	Nitrile
Technical ingredients	Nitrile NBR, Titanium Dioxide, Zinc oxide, Sulfur, ZDBC, Ammonium, Antioxidant, Calcium stearate, Calcium Nitrate, Chlorine
Length/depth	240 mm
Width	115 mm
Thickness	0,07 +/- 0,02 mm
Certifications	CE, Food contact materials, CAT III, MD
CE Category (Personal Protective Equipment)	CAT III
CE Class (Medical Devices)	Class I
Product or test standards	AQL 1.0, EN ISO 374-1:2016 Type B KPT, EN 374-5:2016 VIRUS, EN 420, EN 421:2010 Radioactive Contamination, EN 455
Directives, regulations and acts	MDR (EU) 2017/745, (EU) 2016/425, (EC) No 1935/2004, BEK nr 681 af 25/05/2020, (EC) No 2023/2006
Because of the products composition it has an expiry date from the production date on:	5 years
Storage Instructions	Store dry, clean and at room temperature.
Product Disposal Instructions	Can be disposed of with normal household waste sorted according to local regulations.
Packaging Disposal Instructions	Can be disposed of with normal household waste sorted according to local regulations.

Instructions for use/application

Examine the gloves for flaws and defects before use.

Packaging data

Unit	Contains	Length	Width	Height	EAN
cII	10 pck	38 cm	25.5 cm	25.5 cm	5703538417276
pck	150 pcs	24 cm	12.4 cm	7.5 cm	5703538417252



Regulation (EU) 2017/745 is a regulation of the European Union on the clinical investigation and sale of medical devices for human use. It repeals Directive 93/42/EEC, which concerns medical devices, and Directive 90/385/EEC, which concerns active implantable medical devices, on 26 May 2021.



The CE mark guarantees that a product is safe to use and complies with all safety precautions. CE stands for Conformité Européenne (European Conformity) and is mainly found on electronic equipment, safety equipment, construction products, and medical devices.



The glass-fork-symbol guarantees that products live up to the demands stated in the European legislation for food contact materials. The symbol is not mandatory but advised on products used for food contact.



Third-party type approval is required for all Category III personal protective equipment (PPE) products. CAT III products will be accompanied by a notified body number for the 3rd party verification of the product's safety.



The product does not contain latex.



The glove has a finger textured surface.



The glove has a medium grip.



The glove is 240 mm long.



The standard specifies the requirements for protective gloves against hazardous chemicals and microorganisms. This part of the standard describes the glove's resistance to bacteria, fungi and viruses.



The standard specifies the requirements for protective gloves against chemicals and microorganisms. Type B has been tested for a permeation time of at least 30 minutes with the chemicals sodium hydroxide 40%, hydrogen peroxide 30% and formaldehyde 37%.



The standard specifies the requirements for protective gloves in relation to general requirements and test methods.



EN 455 consists of four standards that gloves must be tested against in order to be a medical glove. This standard specifies the requirements for disposable medical gloves.

